


Kit - J903 Remote Port (P/N 403457)



KEY	PART NUMBER	DESCRIPTION	Qty
1	840553	Conn-Body/F. 206434-1	1
2	840487	Conn-Pin/M AMP 66506-9	8
3	141607	Conn-Hood AMP 206062-1	1
4	141603	Heathshrink 3/4x2,75"	1
5	404345	Heathshrink 3/8	8cm (2")
6	D3728	Instructions – J903 Repair Kit	1

Connections/Anschlüsse/Connexions/aansluitingen

Standard Connections

Pin	Function
1	+12V
2	+12V
3	Remote Zero
4	Segment Data
5	Clock Data
6	Ground
7	Shield
8	Ground

Please Note: Use the enclosed heat shrink to properly seal off the connector, in order to avoid any entering of moisture

Achtung: Gebrauchen Sie die beigefügte Schrumpfhülse um den Konnektor richtig abzudichten, um so Feuchteindringung vorzubeugen.

Attention: Appliquez la hausse rétractable ci-joint pour étancher le connecteur correctement, pour éviter la pénétration de humidité

Belangrijk: Gebruik de bijgeleverde krimpkous om de connector goed af te dichten om zo het binnendringen van vocht te voorkomen.